



Department of Agriculture
PHILIPPINE FIBER INDUSTRY DEVELOPMENT AUTHORITY

3/F DA-PCAF Bldg., Department of Agriculture, Elliptical Rd., Diliman, Quezon City
Tel. Nos. 721-9813 / 928.8756 to 65 Loc.
Website: www.philfida.da.gov.ph

Supplemental/Bid Bulletin for the Procurement of Calibration/Function Test of Laboratory Equipment

Tuesday, 6 August 2019

Addenda for the Procurement of Calibration/Function Test of Laboratory Equipment (4 Lots) - P 1, 436,000.00 (PhilGEPS Ref. No. 6362662) Lot 1- Calibration/Function Test of Analytical Balance, ph Meter, Oven, Furnace, Water Bath, Vacuum Oven and Top Loading Balance - P 41,500. 00 and Lot 3 – Calibration/Function Test of Pulp and Paper Apparatus/Equipment - P 400,000.00 has been issued to modify or amend the items indicated in the Bidding Documents which were issued earlier

Such addenda shall form part of the integral part of the Bid Documents.

All prospective bidders must comply with the agreed specifications as detailed in the Supplemental/Bid Bulletin below:

1. Bids must be duly received by the BAC Secretariat on or before August 22, 2019 at 9:00 AM.

Bid Opening shall be on August 22, 2019 at 9:30 AM at PhilFIDA Conference Room

2. The specification under Section VII – Technical Specifications are as follows:

- a. Lot 1 - Calibration/Function Test of Analytical Balance, ph Meter, Oven, Furnace, Waterbath, Vacuum Oven and Top Loading Balance

Calibration/Function Test of the following:
- Two (2) Units Analytical Balance, (Ohaus, Sartorius)
- One (1) Unit ph Meter, (Oakton pH 700)
- Four (4) Units Oven, (Jouan-2, Memmert (UN 110), Binder)
- One (1) Unit Furnace, (Vulcan, A-550)
- Three (3) Units Waterbath, (Polyscience-2, Memmert)
- One (1) Unit Vacuum Oven (Cole Parmer)
- Four (4) Units Top Loading Balance, (Shimadzu, Radwag (2), AND)
SCOPE OF WORK:
- Physical check-up of the units
- Functional test for proper operation with a Known standard test for instruments
- Calibration by means of comparison with accepted standards for each equipment
- Cleaning of external of instruments
- Issuance of certificate of calibration and calibration sticker

- b. Lot 3 - Calibration/ Function Test of Pulp and Paper Apparatus/Equipment

Calibration of the following:
1. L & W Canadian Freeness
2. L & W Micrometer
3. L & W Densometer
Function Test of the following:
1. L & W Tear Tester
2. NORAM, Wet Disintegrator
3. L & W Rapid Dryer
SCOPE OF WORK:
- Physical check-up of the units
- Functional test for proper operation with A known standard test of instruments
- Calibration by means of comparison with Accepted standards for each equipment
- Cleaning of external part of in instruments
- Issuance of certificate of calibration and calibration sticker


EDEL M. DONDONILLA
BAC Chairperson