



Republic of the Philippines
Department of Agriculture
PHILIPPINE FIBER INDUSTRY DEVELOPMENT AUTHORITY
3/F DA-PCAF Bldg., Department of Agriculture, Elliptical Rd., Diliman, Quezon City
BIDS & AWARDS COMMITTEE
Email: philfidabac@yahoo.com 273.AGRI (2474) / 928-8756 to 65 Local 2652

**Supplemental/Bid Bulletin for
the Supply, Delivery, Installation and Training of Real Time PCR Cycler
(Research Division – PhP 3,000,000.00)**

October 26, 2020


Addendum for the ***Supply, Delivery, Installation and Training of Real Time PCR Cycler (Research Division – PhP 3,000,000.00)*** (PhilGEPS Ref. No. 7256727) has been issued to modify or amend the item as indicated in the ***Invitation to Bid (ITB) and Public Bidding Document (PBD)*** which was issued earlier.

Such addendum shall form part of the bidding documents.

All prospective bidders must comply with the agreed change of description as detailed in the Supplemental/Bid Bulletin below:

- As specified by the end-user in the Technical Specification, the PCR format shall only be Rotary type with specific carousels on rotors and shall be of a circulated air technology.
- On page 69 of the Technical Specifications, the temperature accuracy which is stated “at most $\pm 0.4^{\circ}\text{C}$ ” can be “ $\pm 0.5^{\circ}\text{C}$ ” and shall be accepted.
- For consumables under accessories, the 0.1ml tubes/capillaries and/or 0.2ml tubes can be offered both, either of two or bundled as long as the consumable is included, either the combination of two, or to choose one of 0.1 or 0.2 tubes.
- The Certificate of Exclusivity for the guaranteed warranty and after sales service should be attached to the bidding documents during the submission of bids.
- Indicate at least one pack or pack of 1,000 on the number of pieces in the packaging of PCR tubes.
- Under accessories, the “Cable serial, cable USB” shall be change to “compatible cable”.
- Certificate of Sole Distributorship shall be accepted. If bidders can provide the specific number of years of the warranty, the Authorized Distributorship shall also be accepted.

- Must write/put “comply” for each line opposite the specifications.
- The sixty (60) day delivery period shall include delivery, inspection, installation and training. Request for an extension of delivery for another thirty (30) days to sixty (60) days is allowed only once and is subject for evaluation and approval.
- Country of origin of the equipment is not allowed to be part of the Technical Specifications but it should pass the international quality standards. Equivalent standard such as ISO, ECE or any validation report is acceptable.
- Only hard copy of the bidding documents shall be accepted. The BAC shall be physically present during the opening of bids.
- Bid Opening shall be on November 4, 2020, 9:30 A.M. instead of November 5, 2020 due to conflict of schedule of the BAC members. Deadline of the submission of bids is on or before November 4, 2020, 09:00 A.M.



RAMON M. BRANZUELA
BAC Chairperson