



SUPPLEMENTAL BID BULLETIN
SBD No. 003-2022

To : All Interested/Accredited Suppliers

Please be informed of the following changes and/or addition to the bid documents on the **“Invitation to Bid for the Supply and Delivery of Brand New Portable Water Bath Incubator”** under PhilGEPS Reference No. 9809253 dated 1 June 2023.


1) Section VII: Technical Specifications

From	To
ABS plastic case with dimension of 200 x 200 x 130 mm (to accommodate stainless pan holding one 96-well plate)	ABS plastic case with dimension of 200 x 200 x 130 mm (to accommodate stainless pan holding one 96-well plate); ABS box is at least 3mm thick
Stainless steel pan with dimension of 176 x 162 x 65 mm, with cover (can accommodate one 96-well plate)	Stainless steel pan with dimension of 176 x 162 x 65 mm, with cover (can accommodate one 96-well plate); Steel pan is 1mm thick
With current cutoff fuse	With current cutoff fuse; Thermal fuse 60 degree Celsius, 10A Current cut off fuse, at least 11A
With display panel for temperature control and setting	With display panel for temperature control and setting; Panel display type Battery indicator: LED or LCD; Thermostat: LED or LCD

Correction:

- The correct title for this procurement shall be **“Invitation to Bid for the Supply and Delivery of Brand New Portable Water Bath Incubator”**. It was erroneously typed without the word **“Water”** in the PhilGEPS posting on the Title and on the Description.

Please be guided accordingly.


Engr. RAMIL B. BARCELONA
BAC Chairperson

June 7, 2023