



Republic of the Philippines
Department of Agriculture
PHILIPPINE FIBER INDUSTRY DEVELOPMENT AUTHORITY
The Bids and Awards Committee

FPUL, BAI Compound, Visayas Ave., Vasra, Quezon City
Email: bac@philfida.da.gov.ph Website: www.philfida.da.gov.ph
Tel/Fax: (02) 8531-7854

SUPPLEMENTAL BID BULLETIN

SBD-009-2024


To: All Interested/Accredited Suppliers;

Please be informed on the following clarifications, changes and / or addition on the bid document for the “Invitation to Bid for the Supply and Delivery of Laboratory Equipment – 8 Lots” published in the PhilGEPS website under Reference Number 11511032:

1. The delivery period for all lots shall be **sixty (60)** calendar days instead of forty-five (45) calendar days;
2. Clarifications in Section VII. Technical Specifications of the following lots shall be:

Lot No.	From	To
Lot No. 4 Supply, Delivery and Setup of Branded and Brand New Polymerase Chain Reaction (PCR) Workstation	1. With 0.3um HEPA filter and UV sterilization system	1. With 0.3um HEPA or ULPA and UV sterilization system
Lot No. 7 Supply, Delivery and Installation of Microbial Biosafety Cabinet	5. With built-in HEPA filter and UV Lamp sterilizer (with at least 253 nanometer emission), with aluminum alloy frame HEPA filter guard	5. With built-in HEPA or ULPA filter and UV Lamp sterilizer (with at least 253 nanometer emission), with aluminum alloy frame HEPA or ULPA filter guard
Lot No. 8 Supply, Delivery, Installation and Orientation of Automated PCR Setup Workstation	4. Built in HEPA / UV system	4. Built in HEPA or ULPA / UV system

Please be guided accordingly.


CONCEPCION D. JOCSON
BAC Chairperson