



REQUEST FOR QUOTATION

RFQ No. 033-2025

The Philippine Fiber Industry Development Authority (PhilFIDA), thru its Bids and Awards Committee (BAC) invites all manufacturers, suppliers and distributors to send their sealed bids/quotations within seven (7) calendar days from publication hereof, for the supply/delivery of the items enumerated below:

| Contracts | Qty. | Unit | Description | Approved Budget (PhP) | Place of Delivery |
|---------------------|------|------|---|-----------------------|--|
| PR No. 25-04-305 | 1 | unit | Supply and Delivery of Benchtop pH Meter <u>Technical Specifications:</u> <ul style="list-style-type: none">• Capable of measuring pH, temperature, and millivolts (mV)• Includes a combination pH/temperature probe for accurate and stable measurements• Auto-calibration function• Equipped with Good Laboratory Practice (GLP)-compliant data management features• USB or equivalent data output interface for external data transfer and documentation• Electrode health indicator (e.g., slope display) visible between calibrations• With large LCD display• Compatible with AC input (e.g., 120V AC 60Hz) and DC output (e.g., 9V/300mA)• With IP54 or higher, splash and dust-resistant• Includes compatible calibration reagents• With warrant of at least 1 year for parts and services A. pH Measurement <ul style="list-style-type: none">• Measuring Range: at least (-2.00 to 19.99) pH• Resolution: at least 0.01 pH• Accuracy: at least 0.01 pH | 45,000.00 | PhilFIDA, 3/F ATI Bldg., Elliptical Road, Diliman, Quezon |


| | | | | | |
|--|--|--|---|------------------|--|
| | | | <ul style="list-style-type: none"> • Temperature compensation: 0 to 100°C <p>B. mV Measurement</p> <ul style="list-style-type: none"> • Measuring range: at least \pm 1999 mV • Resolution: at least 0.01 pH • Accuracy: \pm 0.1% FS <p>C. Temperature Measurement</p> <ul style="list-style-type: none"> • Measuring range: at least 0 to 80°C • Resolution: at least 0.1°C • Accuracy: \pm 0.5°C <p><u>Delivery Period:</u></p> <ul style="list-style-type: none"> • Thirty (30) Calendar Days | | |
| | | | TOTAL PhP | 45,000.00 | |

Interested bidders must be included in the registry of manufacturers/ suppliers/ distributors in the PhilGEPS. Please submit your **Bid/Quotation, PhilGEPS Certificate and Mayor's/Business Permit** in a sealed envelope addressed to Mr. Victor Y. Romano. Information relating to examination, evaluation and comparison of price quotations shall be confidential.

Deadline for submission of quotation: **May 23, 2025** at PhilFIDA, Fiber Processing and Utilization Laboratory, BAI Compound, Visayas Ave., Brgy. Vasra, Quezon City.

This document and the Purchase Request can be viewed and downloaded from the Bids and Awards page of the PhilFIDA website (www.philfida.da.gov.ph).

For inquiries, please call BAC Secretariat at Telephone No. 8531-7854 or 09189402419.


JOEL B. BARQUILLA
 Chairperson